
Midmark M11 - v1.0.0
Total Cycles: 5307

Name: RITTER 1
Sterilizer ID: V2178198
02/03/2025 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.2 F 0.0
2:00 68.2 F 0.0
4:00 83.4 F 0.0
6:00 123.3 F 2.1
8:00 153.0 F 4.9
10:00 177.7 F 8.0
12:00 206.5 F 11.7
14:00 211.4 F 3.0
16:00 220.5 F 6.5
18:00 230.6 F 10.2
20:00 231.8 F 8.6
22:00 238.9 F 10.3
24:00 236.1 F 8.3
26:00 246.8 F 13.6
28:00 258.3 F 19.8
30:00 268.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 28.9
2:00 271.6 F 28.2
3:00 272.7 F 29.0
4:00 271.8 F 28.2

Min 271.6 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.9 F 0.0
5:00 209.7 F 0.0
10:00 195.4 F 0.0
15:00 197.6 F 0.0
20:00 177.4 F 0.0
25:00 166.8 F 0.0
30:00 154.9 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 31:13
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:09:23
CYCLE COMPLETE

02/03/2025 10:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____