
Midmark M11 - v1.0.0
Total Cycles: 3555

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2023 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.8 F 0.0
2:00 184.1 F 0.7
4:00 182.6 F 3.5
6:00 202.9 F 7.9
8:00 218.4 F 13.1
10:00 226.1 F 8.2
12:00 238.7 F 13.3
14:00 238.9 F 10.6
16:00 236.5 F 8.5
18:00 250.0 F 15.3
20:00 261.5 F 21.8
22:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.1 F 28.8
2:00 271.5 F 28.6
3:00 272.0 F 28.7
4:00 271.4 F 28.6

Min 271.2 F 28.2
Max 272.3 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 206.2 F 0.0
10:00 224.1 F 0.0
15:00 234.1 F 0.0
20:00 229.3 F 0.0
25:00 222.7 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

07/20/2023 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____