

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4421

Name: RITTER 1  
Sterilizer ID: V2178198  
04/02/2024 10:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
2:00 171.8 F 2.3  
4:00 171.0 F 5.7  
6:00 193.6 F 10.1  
8:00 215.1 F 4.4  
10:00 228.8 F 8.0  
12:00 239.8 F 12.6  
14:00 236.8 F 9.5  
16:00 239.2 F 9.4  
18:00 247.6 F 13.9  
20:00 259.3 F 20.3  
22:00 270.0 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.6 F 28.7  
2:00 271.3 F 28.2  
3:00 272.6 F 28.8  
4:00 271.5 F 28.1

Min 271.3 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.7 F 0.0  
5:00 201.4 F 0.0  
10:00 192.4 F 0.0  
15:00 187.8 F 0.0  
20:00 184.1 F 0.0  
25:00 182.3 F 0.0  
30:00 180.7 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 22:53  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:01:29  
CYCLE COMPLETE

04/02/2024 11:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_