

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3959

Name: RITTER2  
Sterilizer ID: V2313228  
09/19/2024 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 95.4 F 0.0  
2:00 97.2 F 0.0  
4:00 109.9 F 0.1  
6:00 150.8 F 1.7  
8:00 175.2 F 4.8  
10:00 201.6 F 8.6  
12:00 208.4 F 12.9  
14:00 221.9 F 6.5  
16:00 233.0 F 10.8  
18:00 231.6 F 7.8  
20:00 242.8 F 13.1  
22:00 242.2 F 11.9  
24:00 254.1 F 18.0  
26:00 264.8 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.0 F 29.8  
2:00 271.8 F 29.2  
3:00 271.6 F 29.7  
4:00 272.1 F 29.4

Min 271.0 F 28.9  
Max 272.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.1 F 0.0  
5:00 194.2 F 0.0  
10:00 207.0 F 0.0  
15:00 197.5 F 0.0  
20:00 200.4 F 0.0  
25:00 197.6 F 0.0  
30:00 199.6 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 27:50  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:06:18  
CYCLE COMPLETE

09/19/2024 02:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_