
Midmark M11 - v1.0.5
Total Cycles: 1854

Name: RITTER2
Sterilizer ID: V2313228
12/19/2022 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.3 F 0.0
2:00 209.6 F 4.2
4:00 214.2 F 10.1
6:00 224.0 F 6.1
8:00 236.7 F 13.0
10:00 239.6 F 11.5
12:00 240.7 F 11.0
14:00 255.1 F 18.6
16:00 267.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.6
2:00 271.5 F 29.0
3:00 272.6 F 29.7
4:00 271.9 F 29.1

Min 271.4 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 224.8 F 0.0
10:00 230.1 F 0.0
15:00 229.4 F 0.0
20:00 227.8 F 0.0
25:00 227.9 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 2:48
HEATING: 17:16
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:12
CYCLE COMPLETE

12/19/2022 12:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____