
Midmark M11 - v1.0.5
Total Cycles: 2955

Name: RITTER2
Sterilizer ID: V2313228
10/02/2023 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.6 F 0.0
2:00 178.7 F 2.4
4:00 189.7 F 7.2
6:00 211.1 F 13.1
8:00 223.4 F 8.5
10:00 234.5 F 10.8
12:00 240.4 F 12.0
14:00 240.1 F 10.8
16:00 253.7 F 17.6
18:00 265.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.4 F 29.7
2:00 271.8 F 28.8
3:00 272.5 F 29.7
4:00 272.1 F 29.0

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 205.1 F 0.0
10:00 201.0 F 0.0
15:00 199.6 F 0.0
20:00 197.0 F 0.0
25:00 193.6 F 0.0
30:00 192.0 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

10/02/2023 10:43 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____