

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1169

Name: RITTER 1  
Sterilizer ID: V2178198  
09/14/2021 10:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
2:00 191.9 F 3.0  
4:00 202.4 F 7.9  
6:00 220.1 F 13.0  
8:00 227.8 F 8.0  
10:00 239.7 F 13.3  
12:00 239.7 F 10.9  
14:00 237.7 F 9.4  
16:00 251.5 F 16.0  
18:00 262.6 F 22.5  
20:00 272.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.5 F 28.8  
2:00 271.7 F 28.4  
3:00 272.5 F 28.8  
4:00 271.6 F 28.2

Min 271.5 F 28.2  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 226.2 F 0.0  
10:00 231.7 F 0.0  
15:00 221.4 F 0.0  
20:00 220.2 F 0.0  
25:00 223.3 F 0.0  
30:00 215.8 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 20:16  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:38  
CYCLE COMPLETE

09/14/2021 11:39 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01