

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2498

Name: RITTER2  
Sterilizer ID: V2313228  
06/12/2023 09:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
2:00 191.9 F 2.6  
4:00 211.5 F 7.3  
6:00 228.8 F 12.8  
8:00 225.6 F 8.8  
10:00 237.0 F 10.1  
12:00 242.7 F 12.4  
14:00 240.8 F 11.2  
16:00 254.2 F 17.9  
18:00 265.5 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.0  
1:00 272.1 F 29.7  
2:00 272.9 F 29.8  
3:00 272.3 F 30.1  
4:00 272.9 F 29.9  
  
Min 271.0 F 28.9  
Max 273.2 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.9 F 0.0  
5:00 224.7 F 0.0  
10:00 244.3 F 0.0  
15:00 224.8 F 0.0  
20:00 224.9 F 0.0  
25:00 227.0 F 0.0  
30:00 223.9 F 0.0  
  
Drying Complete

FILLING: 1:56  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:51  
CYCLE COMPLETE

06/12/2023 10:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_