

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3775

Name: RITTER2  
Sterilizer ID: V2313228  
07/12/2024 10:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.2 F 0.0  
2:00 131.2 F 0.8  
4:00 141.8 F 2.7  
6:00 166.6 F 5.5  
8:00 183.4 F 9.1  
10:00 202.7 F 12.8  
12:00 212.5 F 5.9  
14:00 223.0 F 9.8  
16:00 233.1 F 10.1  
18:00 236.6 F 10.6  
20:00 235.8 F 8.7  
22:00 246.5 F 13.9  
24:00 257.0 F 19.5  
26:00 266.7 F 25.7  
28:00 273.2 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.3 F 29.5  
2:00 272.5 F 29.9  
3:00 271.2 F 29.3  
4:00 272.3 F 29.8

Min 271.1 F 29.2  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.6 F 0.0  
5:00 215.5 F 0.0  
10:00 188.4 F 0.0  
15:00 181.8 F 0.0  
20:00 177.3 F 0.0  
25:00 174.9 F 0.0  
30:00 173.3 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 28:17  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:06:58  
CYCLE COMPLETE

07/12/2024 11:21 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_