
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000171
10:38 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:32
PRESSURIZING 4:32
STERILIZING 5:40
136.0 C 223kPa 5:40
136.3 C 224kPa 6:10
136.4 C 225kPa 6:40
136.2 C 225kPa 7:10
136.3 C 225kPa 7:40
136.4 C 226kPa 8:10
136.3 C 225kPa 8:40
136.2 C 224kPa 9:10
136.2 C 224kPa 9:10
VENTING 9:10
AIR DRYING 10:10
CYCLE COMPLETE 25:11

Digital Signature #
BE5EA3CC05A30005
