

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1478

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2022 03:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.1 F 0.0  
2:00 172.0 F 3.3  
4:00 187.5 F 7.6  
6:00 213.1 F 12.8  
8:00 222.8 F 7.6  
10:00 237.4 F 13.0  
12:00 237.2 F 10.7  
14:00 236.4 F 9.3  
16:00 250.8 F 16.3  
18:00 263.1 F 23.4  
20:00 272.0 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.4 F 29.6  
2:00 271.4 F 29.4  
3:00 272.2 F 29.6  
4:00 271.2 F 29.2  
  
Min 271.2 F 29.0  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 226.8 F 0.0  
10:00 233.6 F 0.0  
15:00 227.3 F 0.0  
20:00 223.9 F 0.0  
25:00 221.7 F 0.0  
30:00 219.6 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:34  
CYCLE COMPLETE

09/12/2022 04:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_