

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2841

Name: RITTER2  
Sterilizer ID: V2313228  
09/07/2023 08:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 76.4 F 0.0  
4:00 80.7 F 0.0  
6:00 99.0 F 0.0  
8:00 145.7 F 0.1  
10:00 163.8 F 2.7  
12:00 183.1 F 5.8  
14:00 198.1 F 9.0  
16:00 210.0 F 12.4  
18:00 223.9 F 6.3  
20:00 230.0 F 10.3  
22:00 235.1 F 9.2  
24:00 241.5 F 12.0  
26:00 239.1 F 10.4  
28:00 251.1 F 16.2  
30:00 261.4 F 22.1  
32:00 271.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.8  
1:00 272.1 F 29.5  
2:00 271.9 F 28.9  
3:00 271.8 F 29.4  
4:00 272.1 F 29.2  
  
Min 271.2 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
5:00 216.0 F 0.0  
10:00 211.8 F 0.0  
15:00 216.9 F 0.0  
20:00 215.5 F 0.0  
25:00 216.7 F 0.0  
30:00 217.7 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:36  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:10:52  
CYCLE COMPLETE

09/07/2023 09:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_