
Midmark M11 - v1.0.0
Total Cycles: 5239

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2025 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.5 F 0.0
2:00 156.7 F 0.3
4:00 161.7 F 2.7
6:00 189.5 F 6.0
8:00 200.5 F 9.8
10:00 221.4 F 10.1
12:00 223.7 F 7.3
14:00 234.9 F 11.6
16:00 232.0 F 7.9
18:00 244.3 F 13.3
20:00 242.2 F 11.6
22:00 253.8 F 17.4
24:00 263.8 F 23.3
26:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.0 F 29.1
2:00 272.0 F 28.8
3:00 271.3 F 28.3
4:00 271.5 F 28.8

Min 271.1 F 28.3
Max 272.2 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 212.1 F 0.0
10:00 209.5 F 0.0
15:00 206.7 F 0.0
20:00 206.6 F 0.0
25:00 212.2 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:21
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:05:10
CYCLE COMPLETE

01/13/2025 01:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____