
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 006914
13:55 07/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:42
STERILIZING 3:55
136.5 C 222kPa 3:55
136.2 C 220kPa 4:25
135.8 C 224kPa 4:55
136.4 C 224kPa 5:25
136.4 C 222kPa 5:55
136.1 C 222kPa 6:25
136.2 C 219kPa 6:55
136.1 C 220kPa 7:25
136.1 C 220kPa 7:25
VENTING 7:25
AIR DRYING 8:15
CASSETTE REMOVED
CYCLE COMPLETE 31:47

Digital Signature #
88326E9A33933E7F
