

-----  
Midmark M11 - v1.0.0  
Total Cycles: 703

Name: MIDMARK  
Sterilizer ID: V2178198  
02/25/2021 12:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.2 F 0.0  
2:00 114.2 F 0.1  
4:00 125.5 F 0.9  
6:00 159.5 F 4.0  
8:00 192.4 F 8.4  
10:00 211.3 F 13.3  
12:00 221.7 F 8.3  
14:00 235.6 F 12.1  
16:00 237.4 F 10.5  
18:00 234.3 F 8.1  
20:00 249.1 F 15.1  
22:00 261.3 F 21.9  
24:00 272.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.3 F 28.9  
2:00 272.7 F 29.4  
3:00 271.3 F 28.6  
4:00 273.0 F 29.1

Min 271.2 F 28.5  
Max 273.1 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.9 F 0.0  
5:00 245.5 F 0.0  
10:00 215.0 F 0.0  
15:00 228.8 F 0.0  
20:00 226.7 F 0.0  
25:00 226.2 F 0.0  
30:00 216.6 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:22  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 01:02:18  
CYCLE COMPLETE

02/25/2021 01:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_