

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2112

Name: RITTER 1  
Sterilizer ID: V2178198  
07/11/2022 10:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.8 F 0.0  
2:00 193.7 F 3.0  
4:00 196.3 F 7.4  
6:00 216.4 F 12.3  
8:00 221.0 F 5.7  
10:00 234.9 F 10.4  
12:00 232.6 F 7.7  
14:00 242.4 F 12.5  
16:00 240.1 F 10.6  
18:00 253.1 F 17.0  
20:00 263.9 F 23.4  
22:00 271.8 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.4 F 29.3  
2:00 272.3 F 28.9  
3:00 271.4 F 28.3  
4:00 271.7 F 28.9

Min 271.1 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 229.3 F 0.0  
10:00 228.8 F 0.0  
15:00 229.2 F 0.0  
20:00 222.4 F 0.0  
25:00 228.4 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 22:12  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:18  
CYCLE COMPLETE

07/11/2022 11:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_