
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003327
13:54 29/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:44
STERILIZING 3:49
136.0 C 220kPa 3:49
136.4 C 222kPa 4:19
136.4 C 223kPa 4:49
136.3 C 222kPa 5:19
136.4 C 224kPa 5:49
136.3 C 224kPa 6:19
136.4 C 222kPa 6:49
136.3 C 224kPa 7:19
136.3 C 224kPa 7:19
VENTING 7:20
AIR DRYING 8:16
CYCLE COMPLETE 9:01

Digital Signature #
801EB4011120C102
