

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2633

Name: RITTER2  
Sterilizer ID: V2313228  
07/11/2023 12:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.5 F 0.0  
2:00 176.4 F 3.7  
4:00 195.9 F 8.5  
6:00 216.3 F 9.8  
8:00 228.3 F 9.5  
10:00 234.3 F 9.1  
12:00 243.7 F 13.5  
14:00 244.1 F 12.9  
16:00 257.0 F 19.9  
18:00 268.8 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.3 F 29.6  
2:00 271.5 F 29.3  
3:00 272.4 F 29.7  
4:00 271.5 F 29.0  
  
Min 271.4 F 29.0  
Max 272.4 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.5 F 0.0  
5:00 223.5 F 0.0  
10:00 224.7 F 0.0  
15:00 222.6 F 0.0  
20:00 219.5 F 0.0  
25:00 218.2 F 0.0  
30:00 216.5 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 19:05  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:27  
CYCLE COMPLETE

07/11/2023 01:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_