
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007351
7:40 10/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:21
STERILIZING 3:16
136.1 C 220kPa 3:16
136.4 C 224kPa 3:46
136.4 C 224kPa 4:16
136.4 C 220kPa 4:46
136.4 C 220kPa 5:16
136.2 C 217kPa 5:46
136.4 C 221kPa 6:16
136.2 C 223kPa 6:46
136.2 C 223kPa 6:46
VENTING 6:46
AIR DRYING 7:35
CYCLE COMPLETE 14:48

Digital Signature #
587552CA098B45B9
