

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2087

Name: RITTER2  
Sterilizer ID: V2313228  
02/23/2023 02:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.6 F 0.0  
2:00 185.9 F 2.5  
4:00 194.0 F 6.8  
6:00 219.6 F 12.3  
8:00 229.4 F 7.8  
10:00 241.4 F 13.5  
12:00 241.2 F 11.9  
14:00 241.3 F 11.2  
16:00 254.9 F 18.4  
18:00 267.0 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.4 F 29.7  
2:00 272.1 F 29.1  
3:00 271.1 F 29.1  
4:00 272.4 F 29.8  
  
Min 271.1 F 28.8  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 213.2 F 0.0  
10:00 230.8 F 0.0  
15:00 230.0 F 0.0  
20:00 227.9 F 0.0  
25:00 229.1 F 0.0  
30:00 228.7 F 0.0  
  
Drying Complete

FILLING: 1:19  
HEATING: 19:26  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:59  
CYCLE COMPLETE

02/23/2023 03:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_