
Midmark M11 - v1.0.0
Total Cycles: 1608

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2022 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 184.4 F 3.8
4:00 189.1 F 8.3
6:00 221.7 F 9.4
8:00 226.6 F 8.3
10:00 240.2 F 13.1
12:00 240.3 F 11.1
14:00 238.3 F 9.6
16:00 252.4 F 16.5
18:00 264.3 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.8
2:00 271.4 F 28.2
3:00 272.5 F 28.9
4:00 271.7 F 28.3

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 221.3 F 0.0
10:00 227.5 F 0.0
15:00 220.9 F 0.0
20:00 235.6 F 0.0
25:00 235.2 F 0.0
30:00 222.6 F 0.0

Drying Complete

FILLING: 1:31
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

02/08/2022 01:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____