

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000392  
7:24 02/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:48  
PRESSURIZING 3:49  
STERILIZING 5:17  
136.4 C 223kPa 5:17  
136.3 C 223kPa 5:47  
136.0 C 221kPa 6:17  
136.3 C 223kPa 6:47  
136.4 C 223kPa 7:17  
136.4 C 224kPa 7:47  
136.3 C 222kPa 8:17  
136.4 C 223kPa 8:47  
136.4 C 223kPa 8:47  
VENTING 8:47  
AIR DRYING 9:50  
CYCLE COMPLETE 38:13

-----  
Digital Signature #  
99830E3DDBB062D9  
-----