
Midmark M11 - v1.0.5
Total Cycles: 3722

Name: RITTER2
Sterilizer ID: V2313228
06/20/2024 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 71.7 F 0.0
4:00 82.0 F 0.1
6:00 111.8 F 1.8
8:00 139.6 F 4.4
10:00 167.7 F 7.6
12:00 213.6 F 10.9
14:00 216.8 F 3.8
16:00 225.0 F 7.5
18:00 233.8 F 11.6
20:00 230.6 F 7.4
22:00 242.9 F 12.7
24:00 240.0 F 10.6
26:00 252.1 F 16.6
28:00 262.7 F 22.8
30:00 272.3 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.0 F 29.1
2:00 272.0 F 29.2
3:00 272.1 F 29.3
4:00 272.3 F 29.5

Min 271.5 F 29.1
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 228.2 F 0.0
10:00 189.3 F 0.0
15:00 181.7 F 0.0
20:00 181.6 F 0.0
25:00 179.4 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:15
CYCLE COMPLETE

06/20/2024 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____