
Midmark M11 - v1.0.0
Total Cycles: 4265

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2024 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.0 F 0.0
2:00 145.1 F 2.4
4:00 162.3 F 5.3
6:00 189.9 F 9.9
8:00 212.2 F 6.0
10:00 219.0 F 7.4
12:00 232.7 F 12.2
14:00 234.5 F 8.7
16:00 242.0 F 10.9
18:00 245.7 F 13.0
20:00 257.8 F 19.4
22:00 268.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 271.9 F 28.4
2:00 272.3 F 28.9
3:00 271.1 F 28.4
4:00 272.0 F 28.5

Min 271.1 F 28.1
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.9 F 0.0
5:00 240.6 F 0.0
10:00 228.8 F 0.0
15:00 210.3 F 0.0
20:00 234.2 F 0.0
25:00 221.2 F 0.0
30:00 192.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:08
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

02/02/2024 04:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____