

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 002524  
8:55 26/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:52  
PRESSURIZING 3:53  
STERILIZING 4:53  
136.0 C 220kPa 4:53  
136.4 C 222kPa 5:23  
136.4 C 223kPa 5:53  
136.0 C 221kPa 6:23  
136.0 C 221kPa 6:53  
136.3 C 223kPa 7:23  
136.4 C 224kPa 7:53  
136.4 C 224kPa 8:23  
136.4 C 224kPa 8:23  
VENTING 8:23  
AIR DRYING 9:19  
CYCLE COMPLETE 9:23

-----  
Digital Signature #  
8DA491E99E202D1D  
-----