
Midmark M11 - v1.0.5
Total Cycles: 1159

Name: RITTER2
Sterilizer ID: V2313228
06/06/2022 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.3 F 0.0
2:00 149.5 F 2.3
4:00 164.6 F 5.8
6:00 188.2 F 10.4
8:00 210.1 F 5.1
10:00 224.0 F 9.4
12:00 231.8 F 9.8
14:00 240.1 F 11.7
16:00 238.1 F 10.0
18:00 251.2 F 16.4
20:00 262.1 F 22.9
22:00 272.7 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 30.0
2:00 272.0 F 29.4
3:00 271.7 F 29.5
4:00 272.4 F 30.0

Min 271.3 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 252.8 F 0.0
10:00 230.7 F 0.0
15:00 224.3 F 0.0
20:00 219.9 F 0.0
25:00 220.1 F 0.0
30:00 214.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:31
CYCLE COMPLETE

06/06/2022 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____