

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4191

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2024 04:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
2:00 183.0 F 2.8  
4:00 182.5 F 6.5  
6:00 196.8 F 10.7  
8:00 216.3 F 5.3  
10:00 223.8 F 8.5  
12:00 235.3 F 12.9  
14:00 234.8 F 9.6  
16:00 238.0 F 9.9  
18:00 244.5 F 13.2  
20:00 255.4 F 18.9  
22:00 265.5 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 271.5 F 29.2  
2:00 272.4 F 29.6  
3:00 271.7 F 29.1  
4:00 271.0 F 28.9

Min 271.0 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 232.6 F 0.0  
10:00 239.4 F 0.0  
15:00 229.3 F 0.0  
20:00 223.3 F 0.0  
25:00 223.5 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 23:54  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:03:10  
CYCLE COMPLETE

01/10/2024 05:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_