
Midmark M11 - v1.0.5
Total Cycles: 1970

Name: RITTER2
Sterilizer ID: V2313228
01/25/2023 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 181.5 F 0.4
4:00 186.9 F 3.4
6:00 207.6 F 8.5
8:00 223.1 F 10.5
10:00 234.1 F 10.5
12:00 233.0 F 8.5
14:00 239.3 F 10.1
16:00 248.9 F 14.9
18:00 261.4 F 22.0
20:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.6 F 29.6
2:00 272.0 F 29.0
3:00 271.3 F 28.7
4:00 272.1 F 29.6

Min 271.2 F 28.5
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 222.2 F 0.0
10:00 234.6 F 0.0
15:00 238.0 F 0.0
20:00 232.6 F 0.0
25:00 222.2 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

01/25/2023 10:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____