
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002644
7:57 02/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:00
STERILIZING 4:06
136.4 C 222kPa 4:06
136.1 C 223kPa 4:36
136.4 C 223kPa 5:06
136.0 C 222kPa 5:36
136.4 C 223kPa 6:06
136.4 C 224kPa 6:36
136.2 C 220kPa 7:06
135.9 C 221kPa 7:36
135.9 C 221kPa 7:36
VENTING 7:36
AIR DRYING 8:25
CYCLE COMPLETE 9:51

Digital Signature #
C271D1E1899083B7
