
Midmark M11 - v1.0.5
Total Cycles: 3087

Name: RITTER2
Sterilizer ID: V2313228
11/06/2023 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.3 F 0.0
2:00 158.2 F 2.0
4:00 180.2 F 5.9
6:00 201.0 F 10.0
8:00 215.9 F 7.0
10:00 224.8 F 8.7
12:00 232.8 F 11.5
14:00 239.0 F 11.4
16:00 238.3 F 10.2
18:00 251.8 F 16.7
20:00 263.2 F 23.4
22:00 272.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 271.8 F 29.6
2:00 272.4 F 29.5
3:00 271.3 F 28.9
4:00 271.8 F 29.7

Min 271.2 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 204.0 F 0.0
10:00 211.3 F 0.0
15:00 206.4 F 0.0
20:00 203.7 F 0.0
25:00 205.6 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

11/06/2023 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____