
Midmark M11 - v1.0.5
Total Cycles: 2923

Name: RITTER2
Sterilizer ID: V2313228
09/25/2023 04:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.0 F 0.0
2:00 181.1 F 3.4
4:00 199.2 F 8.0
6:00 215.7 F 12.4
8:00 224.9 F 7.4
10:00 236.1 F 12.6
12:00 237.0 F 10.6
14:00 237.2 F 9.5
16:00 251.0 F 16.2
18:00 262.3 F 22.8
20:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.4
1:00 272.3 F 29.5
2:00 272.4 F 29.8
3:00 272.4 F 30.0
4:00 271.8 F 29.3

Min 271.7 F 29.3
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 233.2 F 0.0
10:00 231.5 F 0.0
15:00 209.3 F 0.0
20:00 195.1 F 0.0
25:00 193.8 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

09/25/2023 05:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____