

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2278

Name: RITTER2  
Sterilizer ID: V2313228  
04/20/2023 02:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
2:00 181.7 F 3.9  
4:00 199.1 F 9.0  
6:00 217.2 F 8.0  
8:00 230.0 F 10.2  
10:00 233.0 F 8.0  
12:00 241.1 F 11.1  
14:00 247.2 F 14.2  
16:00 260.6 F 21.5  
18:00 272.2 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.7 F 29.5  
2:00 271.7 F 28.7  
3:00 272.2 F 29.6  
4:00 272.5 F 29.3  
  
Min 271.3 F 28.6  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
5:00 214.4 F 0.0  
10:00 232.6 F 0.0  
15:00 232.7 F 0.0  
20:00 225.0 F 0.0  
25:00 220.7 F 0.0  
30:00 220.5 F 0.0  
  
Drying Complete

FILLING: 1:58  
HEATING: 18:29  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:41  
CYCLE COMPLETE

04/20/2023 03:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_