
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000265
11:28 22/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:07
STERILIZING 4:02
136.4 C 221kPa 4:02
136.4 C 224kPa 4:32
136.4 C 224kPa 5:02
136.4 C 225kPa 5:32
136.1 C 224kPa 6:02
136.4 C 224kPa 6:32
136.2 C 223kPa 7:02
136.0 C 219kPa 7:32
136.0 C 219kPa 7:32
VENTING 7:32
AIR DRYING 8:22
CYCLE COMPLETE 8:29

Digital Signature #
7B394CF01C808CA2
