
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001135
16:15 22/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:44
STERILIZING 4:57
136.4 C 220kPa 4:57
136.4 C 226kPa 5:27
136.3 C 224kPa 5:57
136.2 C 224kPa 6:27
136.4 C 226kPa 6:57
136.4 C 226kPa 7:27
136.4 C 224kPa 7:57
136.4 C 225kPa 8:27
136.4 C 225kPa 8:27
VENTING 8:27
AIR DRYING 9:22
CYCLE COMPLETE 15:13

Digital Signature #
9F4D186BEA90DCD5
