
Midmark M11 - v1.0.0
Total Cycles: 4739

Name: RITTER 1
Sterilizer ID: V2178198
07/17/2024 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 71.5 F 0.0
4:00 80.0 F 0.1
6:00 114.5 F 0.6
8:00 145.9 F 3.1
10:00 167.9 F 5.9
12:00 184.7 F 8.9
14:00 204.4 F 12.2
16:00 213.1 F 4.0
18:00 222.6 F 7.7
20:00 232.2 F 11.5
22:00 230.2 F 7.2
24:00 241.7 F 12.6
26:00 239.7 F 10.5
28:00 252.0 F 16.4
30:00 262.4 F 22.3
32:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.8 F 29.4
2:00 271.8 F 28.6
3:00 272.5 F 29.4
4:00 272.4 F 29.1

Min 271.5 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 204.7 F 0.0
10:00 203.3 F 0.0
15:00 215.0 F 0.0
20:00 229.6 F 0.0
25:00 232.3 F 0.0
30:00 224.0 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 33:20
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:11:43
CYCLE COMPLETE

07/17/2024 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____