

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4087

Name: RITTER2  
Sterilizer ID: V2313228  
11/05/2024 02:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.3 F 0.0  
2:00 175.8 F 3.1  
4:00 182.8 F 7.4  
6:00 212.9 F 11.7  
8:00 216.7 F 6.3  
10:00 225.7 F 11.2  
12:00 231.6 F 8.0  
14:00 242.1 F 12.5  
16:00 243.8 F 12.8  
18:00 255.5 F 18.9  
20:00 266.1 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.4 F 29.5  
2:00 271.8 F 29.9  
3:00 271.8 F 29.3  
4:00 272.1 F 29.9  
  
Min 271.2 F 29.2  
Max 272.5 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 218.9 F 0.0  
10:00 213.0 F 0.0  
15:00 211.7 F 0.0  
20:00 209.6 F 0.0  
25:00 211.7 F 0.0  
30:00 211.2 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 21:44  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:01:09  
CYCLE COMPLETE

11/05/2024 03:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_