

-----  
Midmark M11 - v1.0.0  
Total Cycles: 841

Name: RITTER 1  
Sterilizer ID: V2178198  
05/27/2021 08:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.2 F 0.0  
2:00 155.1 F 1.4  
4:00 163.4 F 4.3  
6:00 187.4 F 8.2  
8:00 207.2 F 12.9  
10:00 219.0 F 5.9  
12:00 232.0 F 10.6  
14:00 234.1 F 7.7  
16:00 244.0 F 12.9  
18:00 242.5 F 11.2  
20:00 254.6 F 17.3  
22:00 265.5 F 23.8  
24:00 271.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.1 F 28.4  
2:00 272.5 F 28.7  
3:00 272.2 F 29.0  
4:00 271.4 F 28.4

Min 271.4 F 28.3  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 235.2 F 0.0  
10:00 222.8 F 0.0  
15:00 222.5 F 0.0  
20:00 222.9 F 0.0  
25:00 224.3 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 24:30  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:02:51  
CYCLE COMPLETE

05/27/2021 09:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_