

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1092

Name: RITTER2  
Sterilizer ID: V2313228  
05/11/2022 09:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.6 F 0.0  
2:00 156.1 F 2.2  
4:00 161.2 F 5.5  
6:00 183.7 F 9.9  
8:00 207.6 F 5.4  
10:00 221.4 F 8.8  
12:00 233.8 F 12.0  
14:00 236.9 F 10.9  
16:00 235.8 F 9.1  
18:00 250.1 F 15.7  
20:00 261.2 F 22.2  
22:00 272.0 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.3 F 29.5  
2:00 272.7 F 29.7  
3:00 272.4 F 29.4  
4:00 272.0 F 29.0

Min 271.3 F 28.7  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 205.3 F 0.0  
10:00 219.7 F 0.0  
15:00 226.9 F 0.0  
20:00 229.5 F 0.0  
25:00 225.7 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 22:43  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:06  
CYCLE COMPLETE

05/11/2022 10:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_