

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3572

Name: RITTER 1  
Sterilizer ID: V2178198  
07/24/2023 04:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.3 F 0.0  
2:00 166.8 F 0.3  
4:00 186.1 F 2.4  
6:00 200.7 F 5.7  
8:00 205.5 F 9.5  
10:00 219.6 F 13.4  
12:00 225.4 F 7.3  
14:00 236.2 F 11.7  
16:00 235.0 F 8.7  
18:00 241.6 F 11.1  
20:00 246.1 F 13.3  
22:00 257.7 F 19.5  
24:00 268.1 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 29.0  
2:00 271.7 F 28.3  
3:00 272.0 F 28.9  
4:00 272.3 F 28.8

Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 210.2 F 0.0  
10:00 211.6 F 0.0  
15:00 208.0 F 0.0  
20:00 204.6 F 0.0  
25:00 204.5 F 0.0  
30:00 201.2 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 25:20  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:03:56  
CYCLE COMPLETE

07/24/2023 05:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_