

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2847

Name: RITTER 1  
Sterilizer ID: V2178198  
02/02/2023 10:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.8 F 0.0  
2:00 167.7 F 3.0  
4:00 177.5 F 7.2  
6:00 198.6 F 12.0  
8:00 223.2 F 5.9  
10:00 235.0 F 11.2  
12:00 233.0 F 7.9  
14:00 242.8 F 12.0  
16:00 244.1 F 12.2  
18:00 256.5 F 18.7  
20:00 267.3 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.0 F 29.1  
2:00 273.0 F 29.3  
3:00 272.8 F 29.0  
4:00 272.2 F 29.3  
  
Min 271.0 F 28.2  
Max 273.1 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
5:00 216.3 F 0.0  
10:00 246.6 F 0.0  
15:00 231.7 F 0.0  
20:00 224.4 F 0.0  
25:00 225.6 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 21:32  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:46  
CYCLE COMPLETE

02/02/2023 11:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_