
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001625
12:48 27/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:15
STERILIZING 3:15
136.5 C 220kPa 3:15
136.2 C 225kPa 3:45
136.2 C 224kPa 4:15
136.4 C 226kPa 4:45
136.4 C 226kPa 5:15
136.4 C 225kPa 5:45
136.2 C 224kPa 6:15
136.3 C 224kPa 6:45
136.3 C 224kPa 6:45
VENTING 6:45
AIR DRYING 7:38
CYCLE COMPLETE 8:02

Digital Signature #
73C500A38DD7A200
