
Midmark M11 - v1.0.5
Total Cycles: 2174

Name: RITTER2
Sterilizer ID: V2313228
03/24/2023 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.7 F 0.0
2:00 137.0 F 0.9
4:00 150.1 F 3.4
6:00 182.0 F 7.7
8:00 205.2 F 13.0
10:00 222.4 F 7.5
12:00 234.0 F 12.8
14:00 235.4 F 10.1
16:00 236.1 F 8.6
18:00 249.6 F 15.3
20:00 261.5 F 22.2
22:00 272.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 271.9 F 29.3
2:00 271.2 F 29.2
3:00 272.0 F 29.3
4:00 271.2 F 29.1

Min 271.1 F 28.7
Max 272.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 214.2 F 0.0
10:00 237.7 F 0.0
15:00 226.0 F 0.0
20:00 222.4 F 0.0
25:00 222.7 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

03/24/2023 12:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____