

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2932

Name: RITTER 1  
Sterilizer ID: V2178198  
02/22/2023 03:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
2:00 203.5 F 5.0  
4:00 221.4 F 10.5  
6:00 227.9 F 6.4  
8:00 240.6 F 12.4  
10:00 239.0 F 10.5  
12:00 238.2 F 9.3  
14:00 253.0 F 16.8  
16:00 265.6 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.4 F 28.7  
2:00 271.8 F 28.6  
3:00 272.3 F 28.7  
4:00 271.5 F 28.4  
  
Min 271.5 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.6 F 0.0  
5:00 233.4 F 0.0  
10:00 226.1 F 0.0  
15:00 224.5 F 0.0  
20:00 223.1 F 0.0  
25:00 223.0 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 2:22  
HEATING: 17:37  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:57:15  
CYCLE COMPLETE

02/22/2023 04:14 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_