
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004794
16:27 29/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:38
STERILIZING 3:38
136.5 C 225kPa 3:38
136.3 C 224kPa 4:08
136.4 C 226kPa 4:38
136.3 C 222kPa 5:08
136.4 C 224kPa 5:38
136.3 C 224kPa 6:08
136.4 C 225kPa 6:38
136.3 C 224kPa 7:08
136.3 C 224kPa 7:08
VENTING 7:08
AIR DRYING 8:08
CYCLE COMPLETE 13:45

Digital Signature #
D18674833699C945
