

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1894

Name: RITTER2  
Sterilizer ID: V2313228  
01/04/2023 01:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.9 F 0.0  
2:00 149.6 F 2.4  
4:00 165.5 F 5.8  
6:00 188.8 F 10.5  
8:00 212.9 F 4.8  
10:00 225.1 F 9.4  
12:00 232.4 F 9.7  
14:00 240.4 F 12.0  
16:00 239.6 F 10.6  
18:00 252.3 F 17.3  
20:00 263.9 F 24.2  
22:00 271.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.5  
1:00 272.6 F 30.0  
2:00 271.7 F 29.5  
3:00 272.7 F 30.1  
4:00 271.9 F 29.4

Min 271.6 F 29.3  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.9 F 0.0  
5:00 229.9 F 0.0  
10:00 242.5 F 0.0  
15:00 235.6 F 0.0  
20:00 228.6 F 0.0  
25:00 230.5 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 22:03  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:22  
CYCLE COMPLETE

01/04/2023 02:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_