
Midmark M11 - v1.0.5
Total Cycles: 4221

Name: RITTER2
Sterilizer ID: V2313228
12/17/2024 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.8 F 0.0
2:00 113.0 F 0.4
4:00 126.7 F 2.4
6:00 159.1 F 5.3
8:00 184.3 F 8.7
10:00 202.7 F 13.2
12:00 212.9 F 6.0
14:00 223.4 F 10.2
16:00 232.5 F 8.7
18:00 238.1 F 11.6
20:00 237.0 F 9.7
22:00 249.4 F 15.7
24:00 260.3 F 21.9
26:00 270.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.6 F 30.1
2:00 272.0 F 29.5
3:00 271.3 F 29.4
4:00 272.3 F 30.1

Min 271.2 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 209.4 F 0.0
10:00 190.1 F 0.0
15:00 186.2 F 0.0
20:00 184.4 F 0.0
25:00 184.4 F 0.0
30:00 185.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:56
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:05:38
CYCLE COMPLETE

12/17/2024 12:48 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____