
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.4uS / 1.5ppm
CYCLE NUMBER 003739
13:40 31/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:28
STERILIZING 4:29
136.4 C 224kPa 4:29
136.4 C 223kPa 4:59
136.4 C 224kPa 5:29
136.4 C 224kPa 5:59
136.3 C 224kPa 6:29
136.3 C 224kPa 6:59
136.4 C 225kPa 7:29
136.4 C 223kPa 7:59
136.4 C 223kPa 7:59
VENTING 7:59
AIR DRYING 8:52
CYCLE COMPLETE 8:53

Digital Signature #
A431568082A5759A
