
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000893
15:05 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:44
STERILIZING 3:51
136.0 C 226kPa 3:51
136.4 C 232kPa 4:21
136.4 C 233kPa 4:51
136.0 C 229kPa 5:21
136.4 C 233kPa 5:51
136.2 C 231kPa 6:21
136.4 C 233kPa 6:51
136.4 C 231kPa 7:21
136.4 C 231kPa 7:21
VENTING 7:21
AIR DRYING 8:14
CYCLE COMPLETE 23:06

Digital Signature #
D456267E2AC6D684
