

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002496  
8:49 22/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:49  
PRESSURIZING 3:49  
STERILIZING 5:00  
136.4 C 220kPa 5:00  
136.3 C 224kPa 5:30  
136.3 C 222kPa 6:00  
136.3 C 223kPa 6:30  
136.1 C 221kPa 7:00  
136.4 C 224kPa 7:30  
136.3 C 223kPa 8:00  
136.1 C 221kPa 8:30  
136.1 C 221kPa 8:30  
VENTING 8:30  
AIR DRYING 9:23  
CYCLE COMPLETE 9:27

-----  
Digital Signature #  
C31B72910AD33BB5  
-----