
Midmark M11 - v1.0.0
Total Cycles: 631

Name: MIDMARK
Sterilizer ID: V2178198
01/26/2021 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.0 F 0.0
2:00 115.9 F 0.7
4:00 136.1 F 2.9
6:00 178.7 F 6.8
8:00 198.0 F 11.8
10:00 213.9 F 6.3
12:00 225.8 F 11.5
14:00 234.0 F 8.3
16:00 240.7 F 10.4
18:00 245.2 F 12.8
20:00 257.8 F 19.7
22:00 269.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.6
2:00 271.7 F 29.1
3:00 272.7 F 29.6
4:00 271.5 F 28.9

Min 271.5 F 28.9
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 235.9 F 0.0
10:00 215.4 F 0.0
Dry Cycle Aborted

FILLING: 0:57
HEATING: 23:53
STERILIZING: 4:00
VENTING: 2:56
DRYING: 12:27
TOTAL CYCLE: 00:44:13

01/26/2021 02:39 PM
ACCEPTED: X
REJECTED:
Operator: U01